


PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

C-Star Calibration

Date **March 26, 2014** S/N# **CST-484PR** Pathlength **25cm**

Analog output

V_d **0.060 V**
 V_{air} **4.804 V**
 V_{ref} **4.723 V**

Temperature of calibration water **19.2 °C**
Ambient temperature during calibration **20.6 °C**

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x , in meters): **$Tr = e^{-cx}$**

To determine beam transmittance: **$Tr = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})$**

To determine beam attenuation coefficient: **$c = -1/x * \ln(Tr)$**

V_d Meter output with the beam blocked. This is the offset.

V_{air} Meter output in air with a clear beam path.

V_{ref} Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain V_{ref} .

Ambient temperature: meter temperature in air during the calibration.

V_{sig} Measured signal output of meter.